

Product XC air temperature sensor
Serial number 077F75
Date of calibration 25 September 24
Customer name
Customer address
Customer reference

Calibration certificate


Specification Air temperature sensor accuracy ± 0.20 °C

Measured values and uncertainties of calibration

Time (mins)	Air temperature (°C)		
	Reference	Measured	Error
0	23.53	23.48	-0.05
10	23.52	23.48	-0.04
20	23.52	23.48	-0.04
30	23.53	23.49	-0.04
40	23.52	23.48	-0.04
50	23.53	23.48	-0.05
60	23.54	23.48	-0.05
Mean error			-0.04
Uncertainty (k=2)			0.07

Reference standards	Ref. no.	Certificate no.	Date
XC air temperature sensor	22T156	22T156-240108-00	08 January 24
Test Procedure	PT-50304-01		30 January 09

Laser measurement system accuracy: When a Renishaw XL-80 laser is used with a Renishaw XC-80 compensator and a Renishaw air temperature sensor (*all within specification*) the laser measurement system accuracy in linear measurement mode will be within: ± 0.5 ppm (k=2) - see the system manual for details.

Authorized signature	Signatory	Position	Date
	Dave Wall	Director & General Manager	25 September 2024

This certificate may not be reproduced other than in full, except with the prior written approval of:

Renishaw plc
 Laser & Calibration Products Division
 Bath Road, Woodchester
 Stroud
 Gloucestershire
 GL5 5EY
 United Kingdom
 Tel +44 (0) 1453 524524

Certificate number
077G75-240925-00

L-9908-1128-02

Calibration notes

1. **Lasers (XM, XL, ML, HS and RLU)** are calibrated by comparison to a reference HeNe laser using an optical beat frequency technique. Reference lasers are routinely calibrated against an iodine-stabilised HeNe laser supplied by the National Physical Laboratory (NPL), or by a national standards laboratory. All frequency measurements are taken over a 1 hour period.
2. **Air pressure and relative humidity (RH) sensors are installed in a compensator (XC and RCU).**
The air pressure sensors are calibrated over 650 mbar to 1150 mbar range in a temperature controlled oven by direct comparison with a reference pressure meter. The RH sensors (where fitted) are certified by the manufacturer to be within specification. They are calibrated by comparison of the readings with those from a reference RH meter at a single applied humidity.
3. **Air and material temperature sensors (XC and RCU)** are calibrated by direct comparison with transfer platinum resistance thermometers (PRTs) in a temperature controlled water bath over 0 °C to 40 °C (50 °C for material sensor). The transfer PRTs are routinely calibrated against reference PRTs.
4. **Rotary axis calibrators (XR20)** are calibrated using a HeNe laser angular interferometer.
5. **Ballbar transducers (QC20 and QC10)** are calibrated using a HeNe laser interferometer. The scale factor (QC10 only) is calculated and must be entered into the Renishaw application software prior to use.
6. **Ballbar calibrators** are calibrated by direct comparison with a reference ballbar calibrator (calibrated by a national standards laboratory) using a reference ballbar as a transfer standard. The measured values for the ballbar calibrator must be entered into the Renishaw application software prior to use.
7. **Traceability.** All the reference standards (listed overleaf) used in these calibrations are traceable either directly to major international metrology institutes who have signed the CIPM Mutual Recognition Agreement (e.g. NPL: UK; LNE: France; NIST: USA; PTB: Germany; NMIJ: Japan) or to a national accreditation body (e.g. UKAS: UK; A2LA: USA).
8. **Environment.** The equipment used for calibration is in a facility held between 15 °C and 25 °C.
9. **Uncertainty calculations.** The uncertainty calculations have been carried out according to the European Co-operation for Accreditation document EA-4/02.
10. **Quality accreditation.** All calibrations above are covered by Renishaw's ISO 9001 quality assurance system. The system is audited and certified by an accredited agency.
11. **Re-calibration.** Customers may wish to confirm that systems are performing within published specifications over time. If so, it is recommended that they should be periodically re-calibrated. Please note that compensators and temperature sensors are re-calibrated only at a single applied temperature, air pressure and humidity. Please refer to the appropriate system manual for further details.